

ELECTRICAL RESEARCH AND DEVELOPMENT ASSOCIATION



R-336, TTC Industrial Area, Thane - Belapur Road, MIDC, Rabale, Navi Mumbai - 400 701.

EPABX : + 91 (022) 27606212 / 13 / 14

E-mail : erdarab@erda.org

Web : http://www.erda.org

**TEST REPORT****SHEET NO.:** 1 OF 2

NAME & ADDRESS OF CUSTOMER M/s. Hatco Industries, "Hazari House", 197-201 Cowasji Hormusji Street, Dhobi Talao, Mumbai - 400 002	TEST REPORT NO.: RP-1819-044256	
	DATE : 12/02/2019	
	CUSTOMER REF. NO.: NIL	
	DATED : 07/02/2019	
	DATE OF SAMPLE RECEIPT	DATE OF TESTING
	10/02/2019	11/02/2019
SAMPLE DESCRIPTION AC VOLTAGE DETECTOR MAKE : HTC Instruments MODEL : AC IV RANGE : 12 - 1000V AC ID. NO. : ACVD-01 FREQUENCY : 50Hz	SAMPLE IDENTIFICATION ERDA Sample Code No.: ERDA-00300512 Enclosure : 1) Annexure-I (Sheet 01No.)	
TEST DETAIL AC Voltage Detection at different Voltage Levels.	TEST SPECIFICATION As per Customer's requirement	
REMARKS: The sample conforms to the requirements of the above test as per customer's specification.		
 CHECKED BY	 NITIN DOSHI APPROVED BY	
Note : 1. This report relates only to the particular sample received in good condition for testing at ERDA. 2. This report cannot be reproduced in part under any circumstances. 3. Publication of this report requires prior permission in writing from Director, ERDA. 4. Only the tests asked for by the customer have been carried out. 5. In case of any dispute, Vadodara will be the exclusive jurisdiction & shall be construed as where the cause has arised.		
Caution: ERDA is not responsible for the authenticity of photocopied or reproduced test reports. ERDA provides support to customers for verification of the authenticity of test reports issued by ERDA.		

0025181



ELECTRICAL RESEARCH AND DEVELOPMENT ASSOCIATION

R-336, TTC Industrial Area, Thane - Belapur Road, MIDC, Rabale, Navi Mumbai - 400 701.

EPABX : + 91 (022) 27606212 / 13 / 14

E-mail : erdarab@erda.org

Web : http://www.erda.org

**TEST REPORT NO.:** RP-1819-044256**SHEET NO.:** 2 OF 2**DATE** :12/02/2019**TEST RESULTS :**

AC Voltage detection at different Voltage Levels by placing the tip of the test sample near the voltage source.

(Test procedure was followed as per Customer's requirement)

ATMOSPHERIC CONDITION : Dry bulb Temperature : 26.5 °C
Wet bulb Temperature : 23.1 °C
Atmospheric Pressure : 760 mm of Hg

VOLTAGE DETECTOR :

Sr. No.	Applied Voltage (AC)	Observation	Remarks
1	12V	Light & Sound indicated	Confirms as per customer's requirement
2	250V	Light & Sound indicated	
3	450V	Light & Sound indicated	
4	600V	Light & Sound indicated	
5	1000V	Light & Sound indicated	

CHECKED BY

0025182

ELECTRICAL RESEARCH AND DEVELOPMENT ASSOCIATION

R-336, TTC Industrial Area, Thane - Belapur Road, MIDC, Rabale, Navi Mumbai - 400 701.

EPABX : + 91 (022) 27606212 / 13 / 14

E-mail : erdarab@erda.org

Web : http://www.erda.org



Annexure-I

TEST REPORT NO.: RP-1819-044256

SHEET NO.: 1 OF 1

DATE :12/02/2019

PHOTOGRAPH OF TEST SAMPLE



ERDA-00300512
SMPW-1819-014879

ERDA SAMPLE CODE: ERDA-00300512

0025183

VR

CHECKED BY

